

EU Declaration of Conformity

No: DML 112 Rev. J

Delta Mobrey Ltd

Riverside Business Park, Dogflud Way, Farnham, Surrey, GU9 7SS, United Kingdom
(Tel: +44(0)1252 729140, Fax: +44(0)1252 729168)

Declares under our sole responsibility that the product(s):

Equipment: **Mobrey™ Hydrastep Electronic Gauging System**

Type Numbers: **2468C*** (Controller); 2457****, 2467**** (Water Columns);
459600***, 24678** (Electrodes)**

(Minor variations in design to suit the application and/or mounting requirements are identified by alpha/numeric characters where indicated * above)

Conform to the relevant provisions of the European Directives:

| | | | |
|-------------------|-----|-------------------|------|
| 2014/30/EU | EMC | 2014/34/EU | ATEX |
| 2014/35/EU | LVD | 2014/68/EU | PED |

Conformity Assessment Procedure followed for 2014/68/EU **B + D**

Inspection carried out by:

For 2014/68/EU:

Irish Engineering Services [2820]

RSA House, Dundrum Town Centre, Sandyford Road,
Dublin 16 D16 FC92, Ireland

| | | |
|--|--|----------------|
| EU Type/Design Examination Certificate to 2014/68/EU | IES/CAT000530347-B1-2 | ≤120Bar |
| EU Type/Design Examination Certificate to 2014/68/EU | IES/CAT000530347-B1-1 | ≤210Bar |
| EU Type/Design Examination Certificate to 2014/68/EU | IES/CAT000530347-B1-3 | ≤300Bar |
| EU Type/Design Examination Certificate to 2014/34/EU | RML 17ATEX4129X | |
| | Category II 3 G Ex nA IIC T4 Gc | |

The following Harmonised Standards have been applied:

EN 61326-1:2013 (Class A) EN 61010-1:2010
EN 60079-0: 2012/A11:2013 EN 60079-15:2010

The following Technical Standards and Specifications have been applied:

EN 61326-3-1:2008 ASME B31.1

Inspection carried out by:

For 2014/68/EU:

Irish Engineering Services [2820]

RSA House, Dundrum Town Centre, Sandyford Road,
Dublin 16 D16 FC92, Ireland

Authorised Signatory for the manufacturer:

Signed:



David J. Ross-Hamilton,

Global Approvals Manager

Date of issue: 23-March-21

Place: Slough, UK